


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2038, Rack A, Plate 1, Well C01, 489992
 [CD769DF652C0A448] Jan 09, 2026 11:25:21 Pressure = 81 bar (80 to 82)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.50	1.1%	11828	
2	0.495	F 0.59	0.7%	7548	
3	0.762	F 0.91	17.5%	191641	
4	0.837	F 1.00	59.3%	649879	Acetylated F peak Consistent with F
5	1.008	F 1.20	0.9%	10252	
6	1.218	A 1.03	20.4%	223239	A peak - REVIEW
7	1.390	S 0.89	0.2%	2196	
Total Area:				1096583	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range