


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2031, Rack A, Plate 1, Well A04, 171680-P  
[883783BAEDDF9BF] Dec 30, 2025 08:51:24 Pressure = 78 bar (78 to 80)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.398	F 0.47	3.2%	17735	
2	0.498	F 0.59	1.3%	7027	
3	0.700	F 0.83	16.9%	94108	
4	0.855	F 1.01	43.5%	242396	Consistent with F
5	0.997	F 1.17	3.3%	18129	
6	1.227	A 1.02	32.0%	178179	A peak - REVIEW
Total Area:				557574	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range