


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2179, Rack A, Plate 1, Well G11, 485918
[9C29BA942810EC4A] Jun 18, 2026 15:23:44 Pressure = 70 bar (70 to 72)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.427	F 0.51	0.8%	8090	
2	0.512	F 0.61	0.4%	4335	
3	0.692	F 0.82	13.8%	146852	Acetylated F peak
4	0.843	F 1.00	66.9%	712534	Consistent with F
5	1.192	A 1.00	18.1%	193182	A peak
Total Area:				1064993	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range