


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1953, Rack A, Plate 1, Well D05, 859813
 [D8710F9F9563A448] Oct 21, 2025 11:38:09 Pressure = 66 bar (65 to 67)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.393	F 0.46	1.0%	14180	
2	0.507	F 0.59	0.3%	4417	
3	0.727	F 0.85	17.6%	239514	Acetylated F peak
4	0.843	F 0.98	67.7%	919135	Consistent with F
5	1.230	A 1.00	13.3%	181160	A peak
Total Area:				1358406	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range