


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2073, Rack A, Plate 1, Well D01, 498930
 [9D36645B3459DB95] Feb 13, 2026 13:46:35 Pressure = 52 bar (52 to 54)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.410	F 0.46	1.1%	13620	
2	0.517	F 0.58	0.5%	5982	
3	0.730	F 0.82	15.4%	186157	Acetylated F peak
4	0.878	F 0.99	63.8%	771817	Consistent with F
5	1.255	A 1.00	19.2%	232304	A peak
Total Area:				1209880	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range