


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1957, Rack A, Plate 1, Well A06, 170723
[99711A7F994B846A] Oct 24, 2025 09:44:04 Pressure = 70 bar (69 to 71)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.392	F 0.46	3.4%	11440	
2	0.490	F 0.57	1.9%	6361	
3	0.717	F 0.84	13.4%	44750	Acetylated F peak
4	0.863	F 1.01	49.5%	164979	Consistent with F
5	0.995	F 1.17	3.0%	9941	
6	1.233	A 1.01	28.8%	95841	A peak - REVIEW
Total Area:				533312	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range