


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1850, Rack A, Plate 1, Well D01, 456897
[8D708CC6771D84C0] Jul 02, 2025 11:43:39 Pressure = 61 bar (60 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.400	F 0.46	1.9%	17084	
2	0.488	F 0.56	3.8%	34603	
3	0.722	F 0.83	12.0%	110581	Acetylated F peak
4	0.867	F 1.00	46.0%	422387	Consistent with F
5	0.977	F 1.13	4.0%	36416	
6	1.113	A 0.92	0.7%	6859	
7	1.220	A 1.01	31.6%	289910	A peak - REVIEW
Total Area:				917840	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range