


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2083, Rack A, Plate 1, Well A04, 492138
[CD7769A2D0640448] Feb 27, 2026 10:24:33 Pressure = 55 bar (55 to 57)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.420	F 0.48	1.2%	9533	
2	0.518	F 0.59	1.2%	10025	
3	0.732	F 0.83	14.7%	120489	Acetylated F peak
4	0.878	F 1.00	71.8%	588913	Consistent with F
5	1.270	A 1.01	11.2%	91704	A peak

Total Area: 820664

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range