


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2124, Rack A, Plate 1, Well G12, 173046
[99370CF6526B5397] Apr 16, 2026 13:02:18 Pressure = 62 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.415	F 0.48	1.0%	8116	
2	0.515	F 0.59	0.7%	5282	
3	0.703	F 0.81	10.3%	83091	Acetylated F peak
4	0.855	F 0.98	48.9%	395480	Consistent with F
5	1.003	F 1.15	3.2%	26219	
6	1.242	A 0.99	36.0%	290853	A peak - REVIEW
Total Area:				809041	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range