


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2140, Rack A, Plate 1, Well F10, 496501
[98271B3F9BE9D315] May 06, 2026 13:51:15 Pressure = 63 bar (63 to 65)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.425	F 0.49	1.1%	12623	
2	0.512	F 0.58	0.5%	5211	
3	0.730	F 0.83	16.1%	185132	Acetylated F peak
4	0.848	F 0.97	70.0%	806931	Consistent with F
5	1.233	A 1.01	12.4%	142487	A peak
Total Area:				1152384	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range