


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2182, Rack A, Plate 1, Well D11, 502197
[8C2CFF789165B937] Jun 23, 2026 11:44:06 Pressure = 72 bar (72 to 74)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.428	F 0.50	1.1%	14322	
2	0.510	F 0.60	0.4%	5384	
3	0.705	F 0.83	16.0%	199867	Acetylated F peak
4	0.823	F 0.97	73.3%	914947	Consistent with F
5	1.187	A 1.01	9.2%	114469	A peak
Total Area:				1248989	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range