


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1380, Rack A, Plate 1, Well B11, 27456
 [89C8B64EE17A4242] Apr 17, 2024 11:09:06

Pressure = 84 bar (84 to 84)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.397	F 0.47	1.5%	13209	
2	0.477	F 0.56	1.4%	12158	
3	0.748	F 0.88	14.3%	122897	Acetylated F peak
4	0.855	F 1.01	55.7%	477781	Consistent with F
5	0.983	F 1.16	1.3%	11514	
6	1.222	A 1.02	25.7%	220939	A peak - REVIEW
Total Area:				858498	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range