


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2194, Rack A, Plate 1, Well G07, 493090
[9C69950AE308EC62] Jul 09, 2026 13:00:54

Pressure = 79 bar (79 to 82)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.690	F	0.83	18.7%	407543	Acetylated F peak
2	0.817	F	0.98	72.4%	1574919	Consistent with F
3	1.143	A	0.99	8.9%	193898	A peak
Total Area: 2176360						

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) A1c > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range