


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2185, Rack A, Plate 1, Well E08, 503336  
[892DEA03216DB117] Jun 25, 2026 13:39:46 Pressure = 72 bar (72 to 74)

FA



| PEAK        | RT    | REL RT | % CONC | AREA    | COMMENT           |
|-------------|-------|--------|--------|---------|-------------------|
| 1           | 0.425 | F 0.51 | 0.7%   | 7488    |                   |
| 2           | 0.510 | F 0.61 | 0.3%   | 3919    |                   |
| 3           | 0.682 | F 0.81 | 13.7%  | 154934  | Acetylated F peak |
| 4           | 0.827 | F 0.99 | 69.1%  | 780146  | Consistent with F |
| 5           | 1.193 | A 1.01 | 16.1%  | 182162  | A peak            |
| Total Area: |       |        |        | 1128649 |                   |

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range