


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2068, Rack A, Plate 1, Well C10, 498921  
 [9D27606E6CF9D93D] Feb 10, 2026 11:32:21 Pressure = 94 bar (93 to 97)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.49	0.8%	9028	
2	0.503	F 0.60	0.5%	5733	
3	0.710	F 0.84	13.7%	151426	Acetylated F peak
4	0.830	F 0.98	52.7%	581676	Consistent with F
5	0.998	F 1.18	2.1%	22936	
6	1.198	A 1.02	29.8%	329332	A peak - REVIEW
7	1.382	S 0.89	0.4%	4350	
Total Area:				1104481	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range