


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2146, Rack A, Plate 1, Well F01, 501860
[8D2DF590AA704C48] May 13, 2026 13:46:41 Pressure = 66 bar (65 to 67)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.515	F 0.59	1.9%	22728	
2	0.723	F 0.83	15.1%	185310	Acetylated F peak
3	0.848	F 0.97	68.3%	838457	Consistent with F
4	1.230	A 1.01	14.7%	180237	A peak
Total Area:				1226732	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range