


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2182, Rack A, Plate 1, Well B07, 485934  
[9C38AE7893C7199D] Jun 23, 2026 10:45:29 Pressure = 73 bar (73 to 75)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.518	F 0.62	0.8%	8430	
2	0.692	F 0.83	13.3%	141044	Acetylated F peak
3	0.825	F 0.99	65.6%	697630	Consistent with F
4	1.183	A 1.01	20.3%	216184	A peak - REVIEW
Total Area:				1063288	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range