


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2041, Rack A, Plate 1, Well A09, 488962
[CD6298365BB5DBB5] Jan 15, 2026 11:02:23 Pressure = 83 bar (83 to 85)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.405	F 0.48	1.2%	10689	
2	0.502	F 0.60	0.4%	3863	
3	0.705	F 0.84	16.8%	144925	Acetylated F peak
4	0.847	F 1.01	72.1%	621168	Consistent with F
5	1.202	A 1.01	9.4%	80536	A peak
Total Area:				861181	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range