


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1924, Rack A, Plate 1, Well A01, 455744  
[993542067E8184C2] Oct 01, 2025 10:41:46 Pressure = 63 bar (62 to 64)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.50	1.3%	10699	
2	0.495	F 0.60	1.2%	9764	
3	0.690	F 0.83	13.2%	109223	Acetylated F peak
4	0.838	F 1.01	61.6%	508590	Consistent with F
5	0.960	F 1.15	2.3%	18673	
6	1.192	A 1.00	20.4%	168522	A peak - REVIEW
Total Area:				825471	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range