


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2068, Rack A, Plate 1, Well G10, 172240
[DC67612E647BFB95] Feb 10, 2026 13:11:23 Pressure = 95 bar (95 to 97)

AF



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.402	F	0.48	2.0%	12027	
2	0.493	F	0.59	1.4%	8508	
3	0.687	F	0.82	6.4%	37886	Acetylated F peak
4	0.845	F	1.01	30.8%	183168	Consistent with F
5	1.007	F	1.20	3.4%	20195	
6	1.198	A	1.01	54.6%	324680	A peak - REVIEW
7	1.382	S	0.89	1.4%	8372	3
Total Area: 594836						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range