


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1875, Rack A, Plate 1, Well C06, 169406
[9931602C3BA6D395] Jul 30, 2025 11:56:11 Pressure = 59 bar (59 to 60)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.48	0.7%	11338	
2	0.495	F 0.59	1.4%	22753	
3	0.833	F 0.99	83.0%	1325010	Consistent with F
4	1.228	A 1.01	14.9%	237218	A peak
Total Area:				1596319	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range