


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2149, Rack A, Plate 1, Well F06, 501551
[CC68B11C1ED8ECEA] May 20, 2026 14:22:04 Pressure = 64 bar (64 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.512	F 0.59	3.4%	36572	
2	0.745	F 0.86	17.2%	182380	Acetylated F peak
3	0.850	F 0.98	70.8%	750776	Consistent with F
4	1.223	A 1.02	8.5%	90536	A peak
Total Area: 1060264					

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range