


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2098, Rack A, Plate 1, Well E03, 172658-p
 [DD273A01549BF397] Mar 19, 2026 12:56:58 Pressure = 60 bar (59 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.410	F 0.46	2.7%	36858	
2	0.508	F 0.57	2.0%	26841	
3	0.723	F 0.82	20.2%	273401	
4	0.877	F 0.99	55.3%	750377	Consistent with F
5	1.123	A 0.91	0.2%	2474	
6	1.237	A 1.00	19.6%	266509	A peak
Total Area:				1356460	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range