


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1801, Rack A, Plate 1, Well A11, 455882  
[8925B847A3B8533F] May 21, 2025 10:11:32 Pressure = 60 bar (59 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.48	0.9%	9177	
2	0.498	F 0.59	1.0%	9687	
3	0.702	F 0.83	15.1%	146701	Acetylated F peak
4	0.842	F 1.00	65.9%	639010	Consistent with F
5	1.228	A 1.01	17.0%	164378	A peak
Total Area:				968953	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range