


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2065, Rack A, Plate 1, Well E07, 172145
[886264EACD0E0E6A] Feb 04, 2026 13:29:24 Pressure = 85 bar (85 to 88)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.510	F	0.61	0.9%	12277	
2	0.692	F	0.82	12.7%	178993	Acetylated F peak
3	0.832	F	0.99	76.8%	1083886	Consistent with F
4	1.190	A	1.01	9.7%	136413	A peak
Total Area: 1411569						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range