


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2179, Rack A, Plate 1, Well F12, 486007  
 [CC6CBEC17FE79997] Jun 18, 2026 15:01:30 Pressure = 70 bar (70 to 72)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.430	F 0.51	0.9%	13070	
2	0.510	F 0.61	0.3%	3734	
3	0.693	F 0.83	15.1%	219506	Acetylated F peak
4	0.832	F 0.99	75.1%	1090025	Consistent with F
5	1.188	A 1.00	8.7%	125957	A peak
Total Area:				1452292	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range