


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2149, Rack A, Plate 1, Well C10, 501893
[C93CF152F4D244C0] May 20, 2026 13:15:23 Pressure = 66 bar (65 to 67)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.515	F 0.60	1.9%	24789	
2	0.713	F 0.83	15.7%	209236	Acetylated F peak
3	0.853	F 1.00	68.9%	920190	Consistent with F
4	1.208	A 1.00	13.6%	182234	A peak
Total Area:				1336449	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range