


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2197, Rack A, Plate 1, Well H07, 173976
[C83DD38AE2AACC40] Jul 15, 2026 15:24:51 Pressure = 83 bar (83 to 85)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.690	F	0.83	30.3%	327967	
2	0.830	F	1.00	59.7%	646452	Consistent with F
3	1.162	A	1.01	10.1%	109055	A peak
Total Area: 1083474						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range