


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1880, Rack A, Plate 1, Well A08, 457234
[DC2477A94DF12E4A] Aug 05, 2025 10:12:00 Pressure = 63 bar (63 to 65)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.49	1.7%	19380	
2	0.487	F 0.58	3.3%	36563	
3	0.690	F 0.82	14.9%	165461	Acetylated F peak
4	0.833	F 0.99	52.6%	583722	Consistent with F
5	0.990	F 1.18	2.5%	28254	
6	1.223	A 1.01	24.9%	275707	A peak - REVIEW
Total Area:				1109087	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range