


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2191, Rack A, Plate 1, Well A06, 486277
[9D2CC43FB9AAECEA] Jul 06, 2026 11:50:36 Pressure = 79 bar (79 to 81)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.508	F	0.61	1.1%	13177	
2	0.708	F	0.85	12.7%	150262	Acetylated F peak
3	0.820	F	0.99	75.4%	893215	Consistent with F
4	1.170	A	1.00	10.8%	128004	A peak
Total Area: 1184658						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range