


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2110, Rack A, Plate 1, Well H07, 498234
[987700140981F135] Mar 31, 2026 14:09:43 Pressure = 58 bar (58 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.507	F 0.57	3.9%	40666	
2	0.725	F 0.81	11.8%	121758	Acetylated F peak
3	0.867	F 0.97	62.5%	646327	Consistent with F
4	1.255	A 1.01	21.8%	225826	A peak - REVIEW
Total Area:				1034577	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range