


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2073, Rack A, Plate 1, Well C03, 492236
 [8867600E63AE2C6A] Feb 13, 2026 13:24:20 Pressure = 52 bar (52 to 54)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.415	F 0.48	1.1%	13981	
2	0.520	F 0.60	0.5%	6301	
3	0.725	F 0.83	15.7%	192299	Acetylated F peak
4	0.867	F 0.99	58.6%	716509	Consistent with F
5	1.015	F 1.16	1.3%	15609	
6	1.257	A 1.01	22.8%	279039	A peak - REVIEW
Total Area:				1223738	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range