


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2204, Rack A, Plate 1, Well A12, 500817
[9C2D8DF9453F3B1F] Jul 23, 2026 10:33:20 Pressure = 93 bar (92 to 94)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.527	F	0.64	1.4%	6514	
2	0.693	F	0.84	16.2%	74557	Acetylated F peak
3	0.845	F	1.02	73.7%	338401	Consistent with F
4	1.170	A	1.01	8.7%	39777	A peak
Total Area: 459249						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range