


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1953, Rack A, Plate 1, Well D11, 170680
[CD641ECAC09CFB9D] Oct 21, 2025 11:50:17 Pressure = 65 bar (65 to 67)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.382	F 0.45	2.6%	15463	
2	0.487	F 0.57	1.0%	6007	
3	0.707	F 0.82	4.7%	27644	Acetylated F peak
4	0.857	F 1.00	24.2%	141856	Consistent with F
5	1.013	F 1.18	5.3%	31275	3
6	1.233	A 1.00	62.0%	362764	A peak - REVIEW
Total Area:				585009	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range