


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1960, Rack A, Plate 1, Well C09, 170800
[8C36B5E5F27A24E8] Oct 28, 2025 10:42:53 Pressure = 67 bar (66 to 69)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.393	F 0.46	1.5%	11391	
2	0.500	F 0.58	0.5%	3888	
3	0.718	F 0.83	13.6%	106217	Acetylated F peak
4	0.852	F 0.99	61.2%	476753	Consistent with F
5	0.998	F 1.16	0.8%	6008	
6	1.240	A 1.01	22.4%	174513	A peak - REVIEW
Total Area:				778770	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range