


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2107, Rack A, Plate 1, Well H03, 498849
[8D7301D40056A6E2] Mar 25, 2026 14:09:34 Pressure = 57 bar (56 to 57)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.423	F 0.48	0.8%	11529	
2	0.515	F 0.58	0.9%	13589	
3	0.728	F 0.83	14.3%	210857	Acetylated F peak
4	0.870	F 0.99	63.4%	932079	Consistent with F
5	1.252	A 1.00	20.5%	301775	A peak - REVIEW
Total Area:				1469829	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range