


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2176, Rack A, Plate 1, Well D05, 493888
[8C3DB9142B90E64A] Jun 12, 2026 13:11:44 Pressure = 68 bar (68 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.510	F 0.60	1.4%	19532	
2	0.698	F 0.83	9.4%	126819	Acetylated F peak
3	0.838	F 0.99	58.0%	780881	Consistent with F
4	1.180	A 0.99	31.2%	419810	A peak - REVIEW
Total Area: 1347042					

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range