


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2003, Rack A, Plate 1, Well F06, 484329
[D862EFA4254DF135] Dec 03, 2025 12:23:03 Pressure = 72 bar (72 to 74)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.47	0.6%	7295	
2	0.502	F 0.58	0.4%	4918	
3	0.708	F 0.83	16.1%	205544	Acetylated F peak
4	0.843	F 0.98	70.3%	897875	Consistent with F
5	1.233	A 1.00	12.6%	160782	A peak

Total Area: 1276414

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range