


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1991, Rack A, Plate 1, Well A01, 171100-P  
 [9922F9B1D0E75995] Nov 21, 2025 14:31:39 Pressure = 70 bar (68 to 72)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.388	F 0.20	6.7%	109622	
2	0.537	F 0.27	8.6%	141287	
3	0.932	F 0.48	20.0%	328043	
4	1.965	F 1.00	25.1%	410396	Consistent with F
5	2.732	F 1.39	4.4%	72303	
6	2.992	A 0.79	1.9%	31108	
7	3.265	A 0.86	0.6%	10077	
8	3.750	A 0.99	31.6%	517196	A peak - REVIEW
9	4.272	A 1.13	0.3%	4977	
10	4.492	A 1.19	0.7%	11741	
				Total Area: 1636750	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range