


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1936, Rack A, Plate 1, Well A07, 170491
[98610C889FBC539D] Oct 08, 2025 11:11:21

Pressure = 66 bar (66 to 67)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.408	F	0.49	0.9%	9510	
2	0.495	F	0.60	0.6%	6441	
3	0.685	F	0.83	12.3%	137504	Acetylated F peak
4	0.822	F	0.99	62.9%	701253	Consistent with F
5	1.175	A	1.01	23.3%	259608	A peak - REVIEW
Total Area: 1114316						

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) A1c > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range