


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2054, Rack A, Plate 1, Well B04, 489158
 [D8239ECAC81DF33F] Jan 26, 2026 11:10:12 Pressure = 88 bar (87 to 90)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.49	1.6%	12769	
2	0.495	F 0.59	0.5%	4153	
3	0.693	F 0.83	7.3%	58234	
4	0.833	F 1.00	77.2%	617665	Consistent with F
5	0.988	F 1.18	1.3%	10129	
6	1.202	A 1.02	12.1%	97044	A peak
Total Area:				799994	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range