


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1875, Rack A, Plate 1, Well B11, 169394
[8C34302C3B86F31D] Jul 30, 2025 11:40:01 Pressure = 59 bar (59 to 60)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.48	1.0%	8059	
2	0.493	F 0.59	0.9%	7682	
3	0.700	F 0.83	11.4%	92493	Acetylated F peak
4	0.850	F 1.01	59.3%	482792	Consistent with F
5	1.090	A 0.90	0.3%	2090	
6	1.218	A 1.01	27.1%	220360	A peak - REVIEW
Total Area:				813476	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range