


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1993, Rack A, Plate 1, Well C10, 482228  
[9873FCC7A8EF7917] Nov 25, 2025 11:41:54 Pressure = 70 bar (68 to 71)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.400	F 0.46	1.4%	12171	
2	0.498	F 0.57	1.1%	9327	
3	0.718	F 0.82	16.7%	147431	Acetylated F peak
4	0.853	F 0.97	73.3%	647913	Consistent with F
5	1.235	A 1.01	7.6%	66986	A peak
Total Area:				883828	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range