


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2107, Rack A, Plate 1, Well E07, 498881
 [C822018157017BBF] Mar 25, 2026 13:02:52 Pressure = 57 bar (57 to 58)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.413	F 0.47	1.2%	15739	
2	0.515	F 0.59	0.9%	12504	
3	0.747	F 0.85	17.5%	238873	Acetylated F peak
4	0.863	F 0.98	71.4%	973904	Consistent with F
5	1.255	A 1.01	9.0%	122391	A peak
Total Area:				1363411	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range