


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1842, Rack A, Plate 1, Well E04, 458018  
[C960D21401DD8CC0] Jun 25, 2025 13:02:03 Pressure = 61 bar (61 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.400	F 0.47	1.1%	13896	
2	0.493	F 0.57	1.0%	13004	
3	0.708	F 0.83	12.8%	160986	Acetylated F peak
4	0.835	F 0.97	63.2%	794286	Consistent with F
5	1.218	A 1.01	21.9%	275006	A peak - REVIEW
Total Area:				1257178	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range