


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1988, Rack A, Plate 1, Well A10, 171107
[CD22B9E08F908648] Nov 21, 2025 10:56:17 Pressure = 70 bar (69 to 72)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.392	F 0.46	1.5%	8231	
2	0.492	F 0.58	0.6%	3509	
3	0.707	F 0.83	1.7%	9812	
4	0.767	F 0.90	1.3%	7416	
5	0.855	F 1.00	49.8%	279271	Consistent with F
6	1.028	F 1.20	1.6%	9152	
7	1.243	A 1.02	43.5%	243947	A peak - REVIEW
Total Area:				561338	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range