


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2107, Rack A, Plate 1, Well A03, 498858
[DC7705940AD60448] Mar 25, 2026 11:15:45 Pressure = 57 bar (57 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.418	F 0.48	0.9%	9034	
2	0.518	F 0.59	0.9%	9031	
3	0.728	F 0.83	19.2%	191417	
4	0.862	F 0.99	62.3%	620268	Consistent with F
5	1.022	F 1.17	2.2%	21572	
6	1.263	A 1.02	14.5%	144358	A peak
Total Area:				995680	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range