


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2197, Rack A, Plate 1, Well D01, 486046
[8D3DC29FBD571BBF] Jul 15, 2026 13:31:40 Pressure = 84 bar (84 to 86)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.513	F	0.62	1.7%	15519	
2	0.700	F	0.85	14.4%	128656	Acetylated F peak
3	0.818	F	0.99	62.9%	562840	Consistent with F
4	1.158	A	1.00	21.0%	187849	A peak - REVIEW
Total Area:					894864	

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range