


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2140, Rack A, Plate 1, Well B07, 173244  
[D9234F7B316BDB35] May 06, 2026 12:04:07 Pressure = 64 bar (64 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.418	F 0.49	2.1%	13392	
2	0.500	F 0.59	2.7%	17097	
3	0.858	F 1.01	62.0%	386400	Consistent with F
4	1.242	A 1.02	33.1%	206640	A peak - REVIEW
Total Area:				623529	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range