


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1988, Rack A, Plate 1, Well A07, 171103
[8C23B9E08F920468] Nov 21, 2025 10:50:13 Pressure = 70 bar (70 to 72)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.398	F 0.47	1.2%	8236	
2	0.495	F 0.58	0.9%	5885	
3	0.705	F 0.82	11.5%	75970	Acetylated F peak
4	0.873	F 1.02	57.7%	381679	Consistent with F
5	1.007	F 1.18	1.8%	11700	
6	1.233	A 1.01	27.0%	178409	A peak - REVIEW
Total Area:				661879	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range